FLUKE®

Calibration

3130 Portable Pressure Calibrator

Delivers best-in-class 300 psi field calibrations

Distributor brochure

A complete system for pressure sensor and pressure gauge calibration, on the bench or in the field.

The 3130 provides a complete solution for calibration of pressure transmitters, gauges, switches, and other such devices for pressures up to 300 psi. With its dual method of pressure generation of a built in pump and external supply pressure connection, it's equally simple and efficient to use in either the field or in the instrument shop/cal lab.

Internal pump capable of generating 2 MPa (300 psi)

- Allows for easy, ergonomic usage in the field.

Supply pressure connection and metering up to 2.2 MPa (330 psi)

- Easy operation in the field or in the shop/lab

Internal pressure sensor with 0.025% reading ± 0.01% FS accuracy

- Capable of handling measurement needs for pressure testing up to 2 MPa (300 psi)

Which customers will use the 3130?

The target application for the 3130 is calibration of pressure transmitters and similar devices in either the field or in a shop/laboratory type environment. This application is especially found in a process plant. Example industries include pharmaceutical, chemical, oil and gas, water treatment, and energy generation.

- **Process manufacturers:** Quality Assurance Manager, Quality Control Inspector, Process Engineer, Validation Engineer, Compliance Manager
- **Third party cal labs:** Metrologist, Lab Manager, Calibration Technician
- **Manufacturing:** Manufacturing Engineer, Asset Manager, Test Engineer, Production Manager

What makes the 3130 a great calibrator?

- **Pressure generation:** The 3130 is unmatched in its capability and flexibility by having both an internal pump and external supply pressure connection capable of being used to the full range of the unit.
- **700P Module compatibility:** The Fluke 700P pressure modules are a popular, well accepted solution for performing industrial level pressure calibrations. The 3130 is compatible with the 700P modules, making it an easy, integrated solution for shops that already have 700P modules.
- **Reliability:** The 3130 is provided by Fluke Calibration, the recognized experts in pressure metrology.

FLUKE ®
Calibration

Ordering information (cont.)

Item Number	Item Model	Description	Gross Weight (kg)	Box Dimensions (LxWxH cm)	Country of Origin	UPC Code	ECCN	HTS	U.S. list price
Mainframe									
4317885	3130-G2M	Portable Gas Pressure Calibrator, LC US, EU, UK, Chin/Aus	7.91	40.64x18.42x32.07	US	0 95969 66543 8	EAR99	9026.20.4000	$6500
4334102	3130-G2M/C	Portable Gas Pressure Calibrator, LC Brz, Ity, Swtz, Ind	7.91	40.64x18.42x32.07	US	0 95969 66843 9	EAR99	9026.20.4000	$7100
4366058	3130-G2M-2	Portable Gas Pressure Calibrator, LC Brz, Ity, Swtz, Ind	7.91	40.64x18.42x32.07	US	0 95969 67801 8	EAR99	9026.20.4000	$6500
4366073	3130-G2M-2/C	Portable Gas Pressure Calibrator, LC Brz, Ity, Swtz, Ind w/Acc Cal	7.91	40.64x18.42x32.07	US	0 95969 67802 5	EAR99	9026.20.4000	$7100

Sales and marketing materials

The following sales and marketing materials are available online at **http://us.flukecal.com/3130-Launch**

- Press release
- Advertisement
- Distributor brochure
- Product brochure
- Data sheet
- Product photos
- Sales guide
- Product announcement
- Sales presentation
- Customer presentation
- Distributor presentation
- Catalog copy
- Web banners
- HTML emails

Fluke Calibration. *Precision, performance, confidence.*™

Electrical | RF | Temperature | Pressure | Flow | Software

Fluke Calibration
PO Box 9090,
Everett, WA 98206 U.S.A.

Fluke Europe B.V.
PO Box 1186, 5602 BD
Eindhoven, The Netherlands

For more information call:
In the U.S.A. (877) 355-3225 or Fax (425) 446-5116
In Europe/M-East/Africa +31 (0) 40 2675 200 or Fax +31 (0) 40 2675 222
In Canada (800)-36-FLUKE or Fax (905) 890-6866
From other countries +1 (425) 446-5500 or Fax +1 (425) 446-5116
Web access: http://www.flukecal.com

©2013 Fluke Corporation. Specifications subject to change without notice.
Printed in U.S.A. 6/2013 6000080A_EN

Modification of this document is not permitted without written permission from Fluke Corporation.